

# QUANTOFIX<sup>®</sup> Ascorbic Acid

---

en

## **Method of application:**

Remove only as many test sticks as are required, and reseal the container immediately after use. Do not touch the test paper zone. Dip the test stick into the test solution for approx. 10 seconds. After 30 seconds, compare the test paper zone with the colour scale. In the presence of ascorbic acid the test paper turns green to blue.

## **Storage:**

Avoid exposing the sticks to sunlight and moisture. Store the container below +30 °C in a dry place.